Multifunction Valves

MFV Cycles

Counts the number of times the Multifuntion Valve (MFV) has been opened. A cycle consists of an open and close. The valve transitions from open to closed when the MFV temperature transistion above 100C (TBC). It is considered closed when it transition from about 100C to below 100C (TBC)

MFV On Time

This is the total powered time of the MFV. This is determined by calculating the duration of time when the MFV temperature is above 100C (TBC).

Revision #1
Created 28 March 2026 21:52:27 by Eric Lyness
Updated 28 March 2026 22:07:43 by Eric Lyness